

DATE _____
SH. 1 OF 1
COMPILED BY
G.T.O.

TMC SPECIFICATION NO. S-184

TITLE: TRANS. SUBSTITUTION TFO01 FOR TF104

JOB

APPROVED G.T.O

110-220 volt operation of DVM using TFO01

In those units where a TF104 (T2) is not available at time of manufacture, a TFO01 is to be substituted and to be wired as shown in the accompanying drawings. The TFO01 differs from a TF104 in that its primary was designed for 110 V. service, where as, the TF104 was designed for 110/220 volt operation. However, the practice is not to be recommended.

